

**CITY AND COUNTY OF SAN FRANCISCO PUBLIC HEALTH LABORATORY** 101 Grove Street, Room 419 San Francisco, CA 94102 Tel: (415) 554-2800 Fax: (415) 431-0651 Godfred Masinde, PhD., Lab Director CLIA ID # 05D0643643

## Test Order AFB (Acid Fast) Smear

	Corporation for the presence of said fast bestill in a clinical specimen
Synonym(s)	Screening for the presence of acid-fast bacilli in a clinical specimen
Methodology	Acid Fast Fluorescent Stain (Auramine-Rhodamine) and Microscopy
Acceptable Specimen	Sputum (expectorated or induced); urine (first morning collection)
Type(s) for Testing	-
Transport / Collection	Sputum: sterile collection cup; urine: large sterile plastic container
Medium	
Storage and	Refrigerate at 4°C after collection and deliver to the laboratory within 3 days.
Preservation of	
Specimen	
Minimum Volume	3 – 10 mL (1 mL is acceptable) sputum; >50 mL urine
Required	
Additional Collection	It is ideal to have at least three consecutive sputum specimens obtained, each
Instructions	collected within 24 hours and at least one being an early morning specimen.
	Have the patient rinse mouth with water before collecting the specimen to
	minimize contamination with food particles, mouth wash, or oral drugs which
	might inhibit the growth of mycobacteria.
Additional Required	Label specimen with patient name, date of birth, medical record number and
Information	date of collection; Label requisition with patient name, date of birth, address,
	medical record number, gender, date of collection, submitter name, requesting
	clinician, specimen source and test request.
Send Out?	No
Turnaround Time	24 hours
Testing Restrictions	Rejection Criteria: Leaky specimens; No label on the specimen; Requisition and
C C	specimen labels do not match; Quantity not sufficient for testing; Specimen is
	too old for testing; Specimen was not stored properly
Requisition Form(s)	http://sfcdcp.org/document.html?id=1035
Limitations / Notes /	Recovery of mycobacteria species is dependent upon quality of specimen
Disclaimer	and collection. It is critical to follow the collection procedure to ensure
	•
	inhibitory substances are removed from the mouth.
	· · · · · · · · · · · · · · · · · · ·

Updated 5/5/2017